



Tender Notice

Non- Resident Nepali Association (NRNA) invites interested eligible Vehicle Suppliers to submit sealed bid documents for supplying Blood Collection & Transportation Vehicles (BCTV) built with all Medical Equipment's required to conduct / support Voluntary Blood Donation and Health Services in Mobile Camps.

Required Number of Vehicles: 3

The detailed Specification of the required BCTV is given below.

| S.N. | Equipment & Specification | Remarks (If any) |
|------|---|------------------|
| 1. | Power: Minimum 125 PS | |
| 2. | Torque: Minimum 400 NM | |
| 3. | Wheelbase: Maximum 3500 mm | |
| 4. | Transmission/ Gear Box 6 Forward, 1Reverse | |
| 5. | Split AC (Genset driven) for operation in only blood collection cabin while vehicle is in stationary. | |
| 6. | BCTV with Engine Driven Air Conditioned and an automotive air conditioner fitted on rooftop. | |
| 7. | Adequate storage cabinets & drawers with labels / markings for keeping Blood reagents, surgical disposables, portable medical equipment's. | |
| 8. | Additional separate Cabinets with markings for Registers & Files, Stationary, IEC material- Pamphlets / posters, housekeeping cloth / mops etc. | |
| 9. | SS Wash Basin - 01 No | |
| 10. | Curtains for all windows & Rear | |
| 11. | Foldable Canopy with extra strong Aluminum structure on co-driver side | |
| 12. | Fans (04 nos), Adequate Lighting, including Emergency Lights | |
| 13. | 240 Volts, 15 / 5 A hybrid sockets with switches in adequate numbers | |
| 14. | Dustbins - 02 units | |
| 15. | Standard fire extinguishers of 1 Kg capacity - 02 Nos | |
| 16. | Generator Set - 01 No | |
| 17. | 1.5 KVA Sine Wave Invertor, with 02 batteries | |
| 18. | PA System | |

| | | |
|-----|--|--|
| 19. | Blood collection Monitor: - 02 Nos, Blood collection unit with end to end mixing and control time to give high quality blood, and with automatic check on blood flow collection time & buzzer showing in LCD display | |
| 20. | Blood Donor Couch, motorized with tilting arrangements: - 02 nos. | |
| 21. | Blood bank refrigerator - 1 No, of capacity 100 Liters (at least 50 bags), with temperature range of 2 - 6 deg C, and holdover time of at least 3 hours | |
| 22. | Table top tube sealer: - 1 no, capable of sealing wide range of blood bag tubing (2-6 mm) separation of tube segments after sealing. Produce Hermatic sealing, hence to contamination & hemolysis. | |
| 23. | Multi-function hand held Tube Stripper: - 1 no, capable for stripping, crimping and cutting of blood bag tube after each blood donation. | |
| 24. | BP Apparatus: - 01 nos. | |
| 25. | Digital Haemoglobinometer (portable) -02 with 500 strips each | |
| 26. | Digital Weighing Balance - 01 | |
| 27. | Mammogram Machine-01 | |

The interested and qualified firms/companies should submit their sealed quotation to the NRNA Secretariat within 15 days from the date of this Notice (27 December, 2022 before 5:00 PM) along with below mentioned documents.

- Official Application letter in Letterhead
- Copy of Company Registration Certificate, VAT Registration Certificate and Tax Clearance Certificate of the recent Fiscal Year
- Documents clarifying Specification, Brand, Warranty Period of the vehicle quoted for

Contact:

NRNA Secretariat

Subarna Shumsher Marga, Baluwatar, POB 1189, Kathmandu, Nepal

Tel; 977-1-4426005; 4411530; Email: info@nrna.org URL: www.nrna.org

Office Hours: Sunday- Friday (10:00 AM till 5:00 PM)

Note: NRNA reserves the right to accept or reject any bid.